

Bresle Patch Test



The Bresle Test will measure water-soluble salts and corrosion products on blast-cleaned steel. These compounds are almost colourless and are localized at the lowest point of the rust pits.

If they are not removed prior to painting, chemical reactions can result in blister formation and accumulations of rust that destroy the adhesion between the substrate and the applied protective coating.

Specification

Conductivity Meter Accuracy: ± 2 .

Conductivity Meter range: 0–1999 μ S/cm.

Conductivity Meter resolution: 1 μ S/cm.

Storage: Do not expose the Bresle Patches to any extremes of temperature or daylight.

Shelf Life: The only degeneration on the Bresle Patches is the adhesive if exposed to extremes of temperature.

We would recommend that the Bresle Patches are used within a 12-month period from date of purchase.

Compliance

ISO 8502-6 and ISO 8502-9.

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com

Supply

Supplied in an industrial foam-filled Carrying Case with Bresle Patches (pack of 35), Conductivity Meter, 500ml Deionised Water 500ml, 5ml Syringe with Needle, Calibration Solution (14ml) Conditioning Solution (14ml) and 25ml Beaker.



Ordering

| | |
|-----------|--|
| T01-30215 | Bresle Test. Includes 35 Bresle Patches |
| T01-30221 | Bresle Patches (pack of 35) |
| T01-30222 | Spare Deionised Water (500ml) |
| T01-30223 | Spare Syringes with Needles (pack of 10) |
| T01-30224 | Spare Conductivity Meter Calibration Solution (14ml) |
| T01-30225 | Spare 25ml Beakers (pack of 5) |
| T01-30226 | Spare Conductivity Meter Sensor Measurement Head |

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com

Instructions

Conductivity Meter Conditioning

For first use on a new Conductivity Meter, condition the measuring electrode with 3 to 4 droplets of the Conditioning Solution, ensuring that the Solution is in both sections of the electrode with no air bubbles and allow to sit for approximately 10 minutes, then the

measuring electrode should be rinsed using the syringe with Deionised Water and shaken dry.



If the measuring electrode has not been used for a long period of time, or if the electrode has been left extremely dry, then use this moistening procedure.

Conductivity Meter Cal

Place 3 to 4 droplets of the 1413 μ S/cm Conductivity Solution into the measuring electrode, ensuring that the solution is in both sections of the electrode with no air bubbles. Check the displayed reading which is shown when the smiley face comes on and if this is not 1413 μ S/cm then calibrate as follows:

Press the Cal button, the CAL indicator and a smiley face will flash – the Conductivity Meter will now auto calibrate. When the CAL indicator and smiley face stop flashing the calibration is complete and the instrument will revert to normal measurement mode.

When you have finished calibrating the Conductivity Meter, the measuring electrode should be rinsed using the syringe with Deionised Water and shaken dry.

Replacing Batteries

When the batteries on the Conductivity Meter require replacement, low battery indicator will show on the display.

With the instrument switched off slide the Sensor while lifting the catch located on the rear of the instrument, replace with 2 lithium CR-2032 batteries, ensuring correct polarity.

Instructions

Measurements

Pour approximately 10ml of Deionised Water into the Beaker.

Completely fill the Syringe with the Deionised Water from the Beaker, and then empty the Syringe back into the Beaker.

Using the Syringe, withdraw approximately 1ml of Deionised Water from the Beaker and place 3 to 4 droplets into the measuring electrode on the Conductivity Meter, ensuring that the Deionised Water is in both sections of the electrode with no air bubbles.

Record the conductivity of the Deionised water displayed by the Meter when the smiley face appears.

Take a Bresle Patch and remove the protective paper and the punched-out center foam. Ensure that you only hold the corner of the Patch away from the adhesive near the test chamber when the protective paper is removed.

The adhesive on the Patch is more adherent in warmer temperatures and this can cause the protective paper stick more. If this is the case remove the backing paper by pushing the punched-out centre foam from the elastomer side.

The centre foam will push off the backing paper when pressure is applied.

Press the adhesive side of the patch against the test surface by running the flat of your finger across from one side of the Patch in such a way that the air in the test chamber is pushed out and the minimum amount of air is trapped.

The elastomer on the Patch should concave inwards and touch the steel in the center of the test

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com

Fill the Syringe with 2.5ml of Deionised Water from the Beaker and insert the Syringe needle at an angle of about 30° to the test surface near the outer edge of the Patch so it passes through the adhesive oam body and into the circular test chamber.

Measurements Continued

Inject the Syringe contents ensuring that it wets the entire test surface, then without removing the Syringe needle from the Patch, suck the contents of the Patch back into the Syringe. Repeat until at least 10 injection–sucking cycles have been completed.

At the end of the 10th cycle retrieve the contaminated water from the Patch with the Syringe and place 3 to 4 droplets into the measuring electrode on the Conductivity Meter, ensuring that the Deionised Water is in both sections of the electrode with no air bubbles.

Record the conductivity of the contaminated water displayed by the Meter when the smiley face appears.

Results

Subtract the initial Deionised Water conductivity reading from the contaminated water conductivity reading. The results are shown in $\mu\text{S}/\text{cm}$.

The Conductivity Meter measurements are shown in $\mu\text{S}/\text{cm}$ and no conversion is required for measurements in mg/m^2 .

For measurements $\mu\text{g}/\text{cm}^2$ add a decimal point in front of the last digit so $100\mu\text{S}/\text{cm}$ will be $10.0\mu\text{g}/\text{cm}^2$ or use the conversion table on the following page.

The conversions listed are based on a test area of 1250mm^2 and using a 2.5ml volume of water.

Expression of results are based on section 7 of ISO 8502-9.

Example. The Deionised Water measurement taken is $4\mu\text{S}/\text{cm}$. The contaminated water measurement taken is $54\mu\text{S}/\text{cm}$. The difference is therefore $50\mu\text{S}/\text{cm}$ which is equivalent to $50\text{mg}/\text{m}^2$ or $5.0\mu\text{g}/\text{cm}^2$.

The Deionised Water temperature can be measured by pressing the MEAS button when the water is in the Conductivity Meter measuring electrode. Press the MEAS button again for normal conductivity measurement mode.

Care

When you have finished using the Conductivity Meter, the measuring electrode should be rinsed using the syringe with Deionised Water and shaken dry. Then place a small amount of Deionised Water in the measuring electrode and replace the sensor cap.

Also ensure the Syringe is cleaned to remove any contamination.

Testing Abrasives

ISO 11127-6: Preparation of steel substrates before application of paints and related products. Test methods for non-metallic blast-cleaning abrasives. Part 6: Determination of water-soluble contaminants by conductivity measurement.

The Bresle Test can also be used for testing non-metallic abrasives for water-soluble salts and corrosion products.

Record the conductivity of the Deionised Water using the same procedure under the section Taking Measurements.

Place 100gm of abrasive into a flask and add 100ml of the Deionized Water that you have recorded the conductivity of.

Shake for 5 minutes and allow to stand for 1 hour. If the liquid does not clear, filter by any suitable method.

Using the Syringe, withdraw approximately 1ml of contaminated water from the flask and place 3 to 4 droplets into the measuring electrode on the Conductivity Meter, ensuring that the contaminated water is in both sections of the electrode with no air bubbles

Record the conductivity of the contaminated water displayed by the Meter when the smiley face appears.

Subtract the initial Deionized Water conductivity reading from the contaminated water conductivity reading. Record the results as shown in $\mu\text{S}/\text{cm}$.



The needles on the Syringes in the Bresle Test are blunt. Care must still be taken when carrying out the test.

When using the Syringes ensure the work area is well lit, be aware of people around you and assess any hazards. Ensure the protective cap is placed over the needle after use.

If the Calibration Solution comes into contact with exposed skin, wash with water. If the Solution comes into contact with eyes, rinse the eye immediately and seek medical advice

Conversions

| Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 | Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 |
|----------------------------|---|--------------------------------------|----------------------------|---|--------------------------------------|
| 1 $\mu\text{S}/\text{cm}$ | 0.1 $\mu\text{g}/\text{cm}^2$ | 1 mg/m^2 | 32 $\mu\text{S}/\text{cm}$ | 3.2 $\mu\text{g}/\text{cm}^2$ | 32 mg/m^2 |
| 2 $\mu\text{S}/\text{cm}$ | 0.2 $\mu\text{g}/\text{cm}^2$ | 2 mg/m^2 | 33 $\mu\text{S}/\text{cm}$ | 3.3 $\mu\text{g}/\text{cm}^2$ | 33 mg/m^2 |
| 3 $\mu\text{S}/\text{cm}$ | 0.3 $\mu\text{g}/\text{cm}^2$ | 3 mg/m^2 | 34 $\mu\text{S}/\text{cm}$ | 3.4 $\mu\text{g}/\text{cm}^2$ | 34 mg/m^2 |
| 4 $\mu\text{S}/\text{cm}$ | 0.4 $\mu\text{g}/\text{cm}^2$ | 4 mg/m^2 | 35 $\mu\text{S}/\text{cm}$ | 3.5 $\mu\text{g}/\text{cm}^2$ | 35 mg/m^2 |
| 5 $\mu\text{S}/\text{cm}$ | 0.5 $\mu\text{g}/\text{cm}^2$ | 5 mg/m^2 | 36 $\mu\text{S}/\text{cm}$ | 3.6 $\mu\text{g}/\text{cm}^2$ | 36 mg/m^2 |
| 6 $\mu\text{S}/\text{cm}$ | 0.6 $\mu\text{g}/\text{cm}^2$ | 6 mg/m^2 | 37 $\mu\text{S}/\text{cm}$ | 3.7 $\mu\text{g}/\text{cm}^2$ | 37 mg/m^2 |
| 7 $\mu\text{S}/\text{cm}$ | 0.7 $\mu\text{g}/\text{cm}^2$ | 7 mg/m^2 | 38 $\mu\text{S}/\text{cm}$ | 3.8 $\mu\text{g}/\text{cm}^2$ | 38 mg/m^2 |
| 8 $\mu\text{S}/\text{cm}$ | 0.8 $\mu\text{g}/\text{cm}^2$ | 8 mg/m^2 | 39 $\mu\text{S}/\text{cm}$ | 3.9 $\mu\text{g}/\text{cm}^2$ | 39 mg/m^2 |
| 9 $\mu\text{S}/\text{cm}$ | 0.9 $\mu\text{g}/\text{cm}^2$ | 9 mg/m^2 | 40 $\mu\text{S}/\text{cm}$ | 4.0 $\mu\text{g}/\text{cm}^2$ | 40 mg/m^2 |
| 10 $\mu\text{S}/\text{cm}$ | 1.0 $\mu\text{g}/\text{cm}^2$ | 10 mg/m^2 | 41 $\mu\text{S}/\text{cm}$ | 4.1 $\mu\text{g}/\text{cm}^2$ | 41 mg/m^2 |
| 11 $\mu\text{S}/\text{cm}$ | 1.1 $\mu\text{g}/\text{cm}^2$ | 11 mg/m^2 | 42 $\mu\text{S}/\text{cm}$ | 4.2 $\mu\text{g}/\text{cm}^2$ | 42 mg/m^2 |
| 12 $\mu\text{S}/\text{cm}$ | 1.2 $\mu\text{g}/\text{cm}^2$ | 12 mg/m^2 | 43 $\mu\text{S}/\text{cm}$ | 4.3 $\mu\text{g}/\text{cm}^2$ | 43 mg/m^2 |
| 13 $\mu\text{S}/\text{cm}$ | 1.3 $\mu\text{g}/\text{cm}^2$ | 13 mg/m^2 | 44 $\mu\text{S}/\text{cm}$ | 4.4 $\mu\text{g}/\text{cm}^2$ | 44 mg/m^2 |
| 14 $\mu\text{S}/\text{cm}$ | 1.4 $\mu\text{g}/\text{cm}^2$ | 14 mg/m^2 | 45 $\mu\text{S}/\text{cm}$ | 4.5 $\mu\text{g}/\text{cm}^2$ | 45 mg/m^2 |
| 15 $\mu\text{S}/\text{cm}$ | 1.5 $\mu\text{g}/\text{cm}^2$ | 15 mg/m^2 | 46 $\mu\text{S}/\text{cm}$ | 4.6 $\mu\text{g}/\text{cm}^2$ | 46 mg/m^2 |
| 16 $\mu\text{S}/\text{cm}$ | 1.6 $\mu\text{g}/\text{cm}^2$ | 16 mg/m^2 | 47 $\mu\text{S}/\text{cm}$ | 4.7 $\mu\text{g}/\text{cm}^2$ | 47 mg/m^2 |
| 17 $\mu\text{S}/\text{cm}$ | 1.7 $\mu\text{g}/\text{cm}^2$ | 17 mg/m^2 | 48 $\mu\text{S}/\text{cm}$ | 4.8 $\mu\text{g}/\text{cm}^2$ | 48 mg/m^2 |
| 18 $\mu\text{S}/\text{cm}$ | 1.8 $\mu\text{g}/\text{cm}^2$ | 18 mg/m^2 | 49 $\mu\text{S}/\text{cm}$ | 4.9 $\mu\text{g}/\text{cm}^2$ | 49 mg/m^2 |
| 19 $\mu\text{S}/\text{cm}$ | 1.9 $\mu\text{g}/\text{cm}^2$ | 19 mg/m^2 | 50 $\mu\text{S}/\text{cm}$ | 5.0 $\mu\text{g}/\text{cm}^2$ | 50 mg/m^2 |
| 20 $\mu\text{S}/\text{cm}$ | 2.0 $\mu\text{g}/\text{cm}^2$ | 20 mg/m^2 | 51 $\mu\text{S}/\text{cm}$ | 5.1 $\mu\text{g}/\text{cm}^2$ | 51 mg/m^2 |
| 21 $\mu\text{S}/\text{cm}$ | 2.1 $\mu\text{g}/\text{cm}^2$ | 21 mg/m^2 | 52 $\mu\text{S}/\text{cm}$ | 5.2 $\mu\text{g}/\text{cm}^2$ | 52 mg/m^2 |
| 22 $\mu\text{S}/\text{cm}$ | 2.2 $\mu\text{g}/\text{cm}^2$ | 22 mg/m^2 | 53 $\mu\text{S}/\text{cm}$ | 5.3 $\mu\text{g}/\text{cm}^2$ | 53 mg/m^2 |
| 23 $\mu\text{S}/\text{cm}$ | 2.3 $\mu\text{g}/\text{cm}^2$ | 23 mg/m^2 | 54 $\mu\text{S}/\text{cm}$ | 5.4 $\mu\text{g}/\text{cm}^2$ | 54 mg/m^2 |
| 24 $\mu\text{S}/\text{cm}$ | 2.4 $\mu\text{g}/\text{cm}^2$ | 24 mg/m^2 | 55 $\mu\text{S}/\text{cm}$ | 5.5 $\mu\text{g}/\text{cm}^2$ | 55 mg/m^2 |
| 25 $\mu\text{S}/\text{cm}$ | 2.5 $\mu\text{g}/\text{cm}^2$ | 25 mg/m^2 | 56 $\mu\text{S}/\text{cm}$ | 5.6 $\mu\text{g}/\text{cm}^2$ | 56 mg/m^2 |
| 26 $\mu\text{S}/\text{cm}$ | 2.6 $\mu\text{g}/\text{cm}^2$ | 26 mg/m^2 | 57 $\mu\text{S}/\text{cm}$ | 5.7 $\mu\text{g}/\text{cm}^2$ | 57 mg/m^2 |
| 27 $\mu\text{S}/\text{cm}$ | 2.7 $\mu\text{g}/\text{cm}^2$ | 27 mg/m^2 | 58 $\mu\text{S}/\text{cm}$ | 5.8 $\mu\text{g}/\text{cm}^2$ | 58 mg/m^2 |
| 28 $\mu\text{S}/\text{cm}$ | 2.8 $\mu\text{g}/\text{cm}^2$ | 28 mg/m^2 | 59 $\mu\text{S}/\text{cm}$ | 5.9 $\mu\text{g}/\text{cm}^2$ | 59 mg/m^2 |
| 29 $\mu\text{S}/\text{cm}$ | 2.9 $\mu\text{g}/\text{cm}^2$ | 29 mg/m^2 | 60 $\mu\text{S}/\text{cm}$ | 6.0 $\mu\text{g}/\text{cm}^2$ | 60 mg/m^2 |
| 30 $\mu\text{S}/\text{cm}$ | 3.0 $\mu\text{g}/\text{cm}^2$ | 30 mg/m^2 | 61 $\mu\text{S}/\text{cm}$ | 6.1 $\mu\text{g}/\text{cm}^2$ | 61 mg/m^2 |
| 31 $\mu\text{S}/\text{cm}$ | 3.1 $\mu\text{g}/\text{cm}^2$ | 31 mg/m^2 | 62 $\mu\text{S}/\text{cm}$ | 6.2 $\mu\text{g}/\text{cm}^2$ | 62 mg/m^2 |

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com

Conversions

| Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 | Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 |
|----------------------------|---|--------------------------------------|-----------------------------|---|--------------------------------------|
| 63 $\mu\text{S}/\text{cm}$ | 6.3 $\mu\text{g}/\text{cm}^2$ | 63 mg/m^2 | 94 $\mu\text{S}/\text{cm}$ | 9.4 $\mu\text{g}/\text{cm}^2$ | 94 mg/m^2 |
| 64 $\mu\text{S}/\text{cm}$ | 6.4 $\mu\text{g}/\text{cm}^2$ | 64 mg/m^2 | 95 $\mu\text{S}/\text{cm}$ | 9.5 $\mu\text{g}/\text{cm}^2$ | 95 mg/m^2 |
| 65 $\mu\text{S}/\text{cm}$ | 6.5 $\mu\text{g}/\text{cm}^2$ | 65 mg/m^2 | 96 $\mu\text{S}/\text{cm}$ | 9.6 $\mu\text{g}/\text{cm}^2$ | 96 mg/m^2 |
| 66 $\mu\text{S}/\text{cm}$ | 6.6 $\mu\text{g}/\text{cm}^2$ | 66 mg/m^2 | 97 $\mu\text{S}/\text{cm}$ | 9.7 $\mu\text{g}/\text{cm}^2$ | 97 mg/m^2 |
| 67 $\mu\text{S}/\text{cm}$ | 6.7 $\mu\text{g}/\text{cm}^2$ | 67 mg/m^2 | 98 $\mu\text{S}/\text{cm}$ | 9.8 $\mu\text{g}/\text{cm}^2$ | 98 mg/m^2 |
| 68 $\mu\text{S}/\text{cm}$ | 6.8 $\mu\text{g}/\text{cm}^2$ | 68 mg/m^2 | 99 $\mu\text{S}/\text{cm}$ | 9.9 $\mu\text{g}/\text{cm}^2$ | 99 mg/m^2 |
| 69 $\mu\text{S}/\text{cm}$ | 6.9 $\mu\text{g}/\text{cm}^2$ | 69 mg/m^2 | 100 $\mu\text{S}/\text{cm}$ | 10.0 $\mu\text{g}/\text{cm}^2$ | 100 mg/m^2 |
| 70 $\mu\text{S}/\text{cm}$ | 7.0 $\mu\text{g}/\text{cm}^2$ | 70 mg/m^2 | 101 $\mu\text{S}/\text{cm}$ | 10.1 $\mu\text{g}/\text{cm}^2$ | 101 mg/m^2 |
| 71 $\mu\text{S}/\text{cm}$ | 7.1 $\mu\text{g}/\text{cm}^2$ | 71 mg/m^2 | 102 $\mu\text{S}/\text{cm}$ | 10.2 $\mu\text{g}/\text{cm}^2$ | 102 mg/m^2 |
| 72 $\mu\text{S}/\text{cm}$ | 7.2 $\mu\text{g}/\text{cm}^2$ | 72 mg/m^2 | 103 $\mu\text{S}/\text{cm}$ | 10.3 $\mu\text{g}/\text{cm}^2$ | 103 mg/m^2 |
| 73 $\mu\text{S}/\text{cm}$ | 7.3 $\mu\text{g}/\text{cm}^2$ | 73 mg/m^2 | 104 $\mu\text{S}/\text{cm}$ | 10.4 $\mu\text{g}/\text{cm}^2$ | 104 mg/m^2 |
| 74 $\mu\text{S}/\text{cm}$ | 7.4 $\mu\text{g}/\text{cm}^2$ | 74 mg/m^2 | 105 $\mu\text{S}/\text{cm}$ | 10.5 $\mu\text{g}/\text{cm}^2$ | 105 mg/m^2 |
| 75 $\mu\text{S}/\text{cm}$ | 7.5 $\mu\text{g}/\text{cm}^2$ | 75 mg/m^2 | 106 $\mu\text{S}/\text{cm}$ | 10.6 $\mu\text{g}/\text{cm}^2$ | 106 mg/m^2 |
| 76 $\mu\text{S}/\text{cm}$ | 7.6 $\mu\text{g}/\text{cm}^2$ | 76 mg/m^2 | 107 $\mu\text{S}/\text{cm}$ | 10.7 $\mu\text{g}/\text{cm}^2$ | 107 mg/m^2 |
| 77 $\mu\text{S}/\text{cm}$ | 7.7 $\mu\text{g}/\text{cm}^2$ | 77 mg/m^2 | 108 $\mu\text{S}/\text{cm}$ | 10.8 $\mu\text{g}/\text{cm}^2$ | 108 mg/m^2 |
| 78 $\mu\text{S}/\text{cm}$ | 7.8 $\mu\text{g}/\text{cm}^2$ | 78 mg/m^2 | 109 $\mu\text{S}/\text{cm}$ | 10.9 $\mu\text{g}/\text{cm}^2$ | 109 mg/m^2 |
| 79 $\mu\text{S}/\text{cm}$ | 7.9 $\mu\text{g}/\text{cm}^2$ | 79 mg/m^2 | 110 $\mu\text{S}/\text{cm}$ | 11.0 $\mu\text{g}/\text{cm}^2$ | 110 mg/m^2 |
| 80 $\mu\text{S}/\text{cm}$ | 8.0 $\mu\text{g}/\text{cm}^2$ | 80 mg/m^2 | 111 $\mu\text{S}/\text{cm}$ | 11.1 $\mu\text{g}/\text{cm}^2$ | 111 mg/m^2 |
| 81 $\mu\text{S}/\text{cm}$ | 8.1 $\mu\text{g}/\text{cm}^2$ | 81 mg/m^2 | 112 $\mu\text{S}/\text{cm}$ | 11.2 $\mu\text{g}/\text{cm}^2$ | 112 mg/m^2 |
| 82 $\mu\text{S}/\text{cm}$ | 8.2 $\mu\text{g}/\text{cm}^2$ | 82 mg/m^2 | 113 $\mu\text{S}/\text{cm}$ | 11.3 $\mu\text{g}/\text{cm}^2$ | 113 mg/m^2 |
| 83 $\mu\text{S}/\text{cm}$ | 8.3 $\mu\text{g}/\text{cm}^2$ | 83 mg/m^2 | 114 $\mu\text{S}/\text{cm}$ | 11.4 $\mu\text{g}/\text{cm}^2$ | 114 mg/m^2 |
| 84 $\mu\text{S}/\text{cm}$ | 8.4 $\mu\text{g}/\text{cm}^2$ | 84 mg/m^2 | 115 $\mu\text{S}/\text{cm}$ | 11.5 $\mu\text{g}/\text{cm}^2$ | 115 mg/m^2 |
| 85 $\mu\text{S}/\text{cm}$ | 8.5 $\mu\text{g}/\text{cm}^2$ | 85 mg/m^2 | 116 $\mu\text{S}/\text{cm}$ | 11.6 $\mu\text{g}/\text{cm}^2$ | 116 mg/m^2 |
| 86 $\mu\text{S}/\text{cm}$ | 8.6 $\mu\text{g}/\text{cm}^2$ | 86 mg/m^2 | 117 $\mu\text{S}/\text{cm}$ | 11.7 $\mu\text{g}/\text{cm}^2$ | 117 mg/m^2 |
| 87 $\mu\text{S}/\text{cm}$ | 8.7 $\mu\text{g}/\text{cm}^2$ | 87 mg/m^2 | 118 $\mu\text{S}/\text{cm}$ | 11.8 $\mu\text{g}/\text{cm}^2$ | 118 mg/m^2 |
| 88 $\mu\text{S}/\text{cm}$ | 8.8 $\mu\text{g}/\text{cm}^2$ | 88 mg/m^2 | 119 $\mu\text{S}/\text{cm}$ | 11.9 $\mu\text{g}/\text{cm}^2$ | 119 mg/m^2 |
| 89 $\mu\text{S}/\text{cm}$ | 8.9 $\mu\text{g}/\text{cm}^2$ | 89 mg/m^2 | 120 $\mu\text{S}/\text{cm}$ | 12.0 $\mu\text{g}/\text{cm}^2$ | 120 mg/m^2 |
| 90 $\mu\text{S}/\text{cm}$ | 9.0 $\mu\text{g}/\text{cm}^2$ | 90 mg/m^2 | 121 $\mu\text{S}/\text{cm}$ | 12.1 $\mu\text{g}/\text{cm}^2$ | 121 mg/m^2 |
| 91 $\mu\text{S}/\text{cm}$ | 9.1 $\mu\text{g}/\text{cm}^2$ | 91 mg/m^2 | 122 $\mu\text{S}/\text{cm}$ | 12.2 $\mu\text{g}/\text{cm}^2$ | 122 mg/m^2 |
| 92 $\mu\text{S}/\text{cm}$ | 9.2 $\mu\text{g}/\text{cm}^2$ | 92 mg/m^2 | 123 $\mu\text{S}/\text{cm}$ | 12.3 $\mu\text{g}/\text{cm}^2$ | 123 mg/m^2 |
| 93 $\mu\text{S}/\text{cm}$ | 9.3 $\mu\text{g}/\text{cm}^2$ | 93 mg/m^2 | 124 $\mu\text{S}/\text{cm}$ | 12.4 $\mu\text{g}/\text{cm}^2$ | 124 mg/m^2 |

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com

Conversions

| Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 | Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 |
|-----------------------------|---|--------------------------------------|-----------------------------|---|--------------------------------------|
| 125 $\mu\text{S}/\text{cm}$ | 12.5 $\mu\text{g}/\text{cm}^2$ | 125 mg/m^2 | 156 $\mu\text{S}/\text{cm}$ | 15.6 $\mu\text{g}/\text{cm}^2$ | 156 mg/m^2 |
| 126 $\mu\text{S}/\text{cm}$ | 12.6 $\mu\text{g}/\text{cm}^2$ | 126 mg/m^2 | 157 $\mu\text{S}/\text{cm}$ | 15.7 $\mu\text{g}/\text{cm}^2$ | 157 mg/m^2 |
| 127 $\mu\text{S}/\text{cm}$ | 12.7 $\mu\text{g}/\text{cm}^2$ | 127 mg/m^2 | 158 $\mu\text{S}/\text{cm}$ | 15.8 $\mu\text{g}/\text{cm}^2$ | 158 mg/m^2 |
| 128 $\mu\text{S}/\text{cm}$ | 12.8 $\mu\text{g}/\text{cm}^2$ | 128 mg/m^2 | 159 $\mu\text{S}/\text{cm}$ | 15.9 $\mu\text{g}/\text{cm}^2$ | 159 mg/m^2 |
| 129 $\mu\text{S}/\text{cm}$ | 12.9 $\mu\text{g}/\text{cm}^2$ | 129 mg/m^2 | 160 $\mu\text{S}/\text{cm}$ | 16.0 $\mu\text{g}/\text{cm}^2$ | 160 mg/m^2 |
| 130 $\mu\text{S}/\text{cm}$ | 13.0 $\mu\text{g}/\text{cm}^2$ | 130 mg/m^2 | 161 $\mu\text{S}/\text{cm}$ | 16.1 $\mu\text{g}/\text{cm}^2$ | 161 mg/m^2 |
| 131 $\mu\text{S}/\text{cm}$ | 13.1 $\mu\text{g}/\text{cm}^2$ | 131 mg/m^2 | 162 $\mu\text{S}/\text{cm}$ | 16.2 $\mu\text{g}/\text{cm}^2$ | 162 mg/m^2 |
| 132 $\mu\text{S}/\text{cm}$ | 13.2 $\mu\text{g}/\text{cm}^2$ | 132 mg/m^2 | 163 $\mu\text{S}/\text{cm}$ | 16.3 $\mu\text{g}/\text{cm}^2$ | 163 mg/m^2 |
| 133 $\mu\text{S}/\text{cm}$ | 13.3 $\mu\text{g}/\text{cm}^2$ | 133 mg/m^2 | 164 $\mu\text{S}/\text{cm}$ | 16.4 $\mu\text{g}/\text{cm}^2$ | 164 mg/m^2 |
| 134 $\mu\text{S}/\text{cm}$ | 13.4 $\mu\text{g}/\text{cm}^2$ | 134 mg/m^2 | 165 $\mu\text{S}/\text{cm}$ | 16.5 $\mu\text{g}/\text{cm}^2$ | 165 mg/m^2 |
| 135 $\mu\text{S}/\text{cm}$ | 13.5 $\mu\text{g}/\text{cm}^2$ | 135 mg/m^2 | 166 $\mu\text{S}/\text{cm}$ | 16.6 $\mu\text{g}/\text{cm}^2$ | 166 mg/m^2 |
| 136 $\mu\text{S}/\text{cm}$ | 13.6 $\mu\text{g}/\text{cm}^2$ | 136 mg/m^2 | 167 $\mu\text{S}/\text{cm}$ | 16.7 $\mu\text{g}/\text{cm}^2$ | 167 mg/m^2 |
| 137 $\mu\text{S}/\text{cm}$ | 13.7 $\mu\text{g}/\text{cm}^2$ | 137 mg/m^2 | 168 $\mu\text{S}/\text{cm}$ | 16.8 $\mu\text{g}/\text{cm}^2$ | 168 mg/m^2 |
| 138 $\mu\text{S}/\text{cm}$ | 13.8 $\mu\text{g}/\text{cm}^2$ | 138 mg/m^2 | 169 $\mu\text{S}/\text{cm}$ | 16.9 $\mu\text{g}/\text{cm}^2$ | 169 mg/m^2 |
| 139 $\mu\text{S}/\text{cm}$ | 13.9 $\mu\text{g}/\text{cm}^2$ | 139 mg/m^2 | 170 $\mu\text{S}/\text{cm}$ | 17.0 $\mu\text{g}/\text{cm}^2$ | 170 mg/m^2 |
| 140 $\mu\text{S}/\text{cm}$ | 14.0 $\mu\text{g}/\text{cm}^2$ | 140 mg/m^2 | 171 $\mu\text{S}/\text{cm}$ | 17.1 $\mu\text{g}/\text{cm}^2$ | 171 mg/m^2 |
| 141 $\mu\text{S}/\text{cm}$ | 14.1 $\mu\text{g}/\text{cm}^2$ | 141 mg/m^2 | 172 $\mu\text{S}/\text{cm}$ | 17.2 $\mu\text{g}/\text{cm}^2$ | 172 mg/m^2 |
| 142 $\mu\text{S}/\text{cm}$ | 14.2 $\mu\text{g}/\text{cm}^2$ | 142 mg/m^2 | 173 $\mu\text{S}/\text{cm}$ | 17.3 $\mu\text{g}/\text{cm}^2$ | 173 mg/m^2 |
| 143 $\mu\text{S}/\text{cm}$ | 14.3 $\mu\text{g}/\text{cm}^2$ | 143 mg/m^2 | 174 $\mu\text{S}/\text{cm}$ | 17.4 $\mu\text{g}/\text{cm}^2$ | 174 mg/m^2 |
| 144 $\mu\text{S}/\text{cm}$ | 14.4 $\mu\text{g}/\text{cm}^2$ | 144 mg/m^2 | 175 $\mu\text{S}/\text{cm}$ | 17.5 $\mu\text{g}/\text{cm}^2$ | 175 mg/m^2 |
| 145 $\mu\text{S}/\text{cm}$ | 14.5 $\mu\text{g}/\text{cm}^2$ | 145 mg/m^2 | 176 $\mu\text{S}/\text{cm}$ | 17.6 $\mu\text{g}/\text{cm}^2$ | 176 mg/m^2 |
| 146 $\mu\text{S}/\text{cm}$ | 14.6 $\mu\text{g}/\text{cm}^2$ | 146 mg/m^2 | 177 $\mu\text{S}/\text{cm}$ | 17.7 $\mu\text{g}/\text{cm}^2$ | 177 mg/m^2 |
| 147 $\mu\text{S}/\text{cm}$ | 14.7 $\mu\text{g}/\text{cm}^2$ | 147 mg/m^2 | 178 $\mu\text{S}/\text{cm}$ | 17.8 $\mu\text{g}/\text{cm}^2$ | 178 mg/m^2 |
| 148 $\mu\text{S}/\text{cm}$ | 14.8 $\mu\text{g}/\text{cm}^2$ | 148 mg/m^2 | 179 $\mu\text{S}/\text{cm}$ | 17.9 $\mu\text{g}/\text{cm}^2$ | 179 mg/m^2 |
| 149 $\mu\text{S}/\text{cm}$ | 14.9 $\mu\text{g}/\text{cm}^2$ | 149 mg/m^2 | 180 $\mu\text{S}/\text{cm}$ | 18.0 $\mu\text{g}/\text{cm}^2$ | 180 mg/m^2 |
| 150 $\mu\text{S}/\text{cm}$ | 15.0 $\mu\text{g}/\text{cm}^2$ | 150 mg/m^2 | 181 $\mu\text{S}/\text{cm}$ | 18.1 $\mu\text{g}/\text{cm}^2$ | 181 mg/m^2 |
| 151 $\mu\text{S}/\text{cm}$ | 15.1 $\mu\text{g}/\text{cm}^2$ | 151 mg/m^2 | 182 $\mu\text{S}/\text{cm}$ | 18.2 $\mu\text{g}/\text{cm}^2$ | 182 mg/m^2 |
| 152 $\mu\text{S}/\text{cm}$ | 15.2 $\mu\text{g}/\text{cm}^2$ | 152 mg/m^2 | 183 $\mu\text{S}/\text{cm}$ | 18.3 $\mu\text{g}/\text{cm}^2$ | 183 mg/m^2 |
| 153 $\mu\text{S}/\text{cm}$ | 15.3 $\mu\text{g}/\text{cm}^2$ | 153 mg/m^2 | 184 $\mu\text{S}/\text{cm}$ | 18.4 $\mu\text{g}/\text{cm}^2$ | 184 mg/m^2 |
| 154 $\mu\text{S}/\text{cm}$ | 15.4 $\mu\text{g}/\text{cm}^2$ | 154 mg/m^2 | 185 $\mu\text{S}/\text{cm}$ | 18.5 $\mu\text{g}/\text{cm}^2$ | 185 mg/m^2 |
| 155 $\mu\text{S}/\text{cm}$ | 15.5 $\mu\text{g}/\text{cm}^2$ | 155 mg/m^2 | 186 $\mu\text{S}/\text{cm}$ | 18.6 $\mu\text{g}/\text{cm}^2$ | 186 mg/m^2 |

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com

Conversions

| Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 | Results | Conversion $\mu\text{g}/\text{cm}^2$ | Conversion mg/m^2 |
|-----------------------------|---|--------------------------------------|-----------------------------|---|--------------------------------------|
| 187 $\mu\text{S}/\text{cm}$ | 18.7 $\mu\text{g}/\text{cm}^2$ | 187 mg/m^2 | 218 $\mu\text{S}/\text{cm}$ | 21.8 $\mu\text{g}/\text{cm}^2$ | 218 mg/m^2 |
| 188 $\mu\text{S}/\text{cm}$ | 18.8 $\mu\text{g}/\text{cm}^2$ | 188 mg/m^2 | 219 $\mu\text{S}/\text{cm}$ | 21.9 $\mu\text{g}/\text{cm}^2$ | 219 mg/m^2 |
| 189 $\mu\text{S}/\text{cm}$ | 18.9 $\mu\text{g}/\text{cm}^2$ | 189 mg/m^2 | 220 $\mu\text{S}/\text{cm}$ | 22.0 $\mu\text{g}/\text{cm}^2$ | 220 mg/m^2 |
| 190 $\mu\text{S}/\text{cm}$ | 19.0 $\mu\text{g}/\text{cm}^2$ | 190 mg/m^2 | 221 $\mu\text{S}/\text{cm}$ | 22.1 $\mu\text{g}/\text{cm}^2$ | 221 mg/m^2 |
| 191 $\mu\text{S}/\text{cm}$ | 19.1 $\mu\text{g}/\text{cm}^2$ | 191 mg/m^2 | 222 $\mu\text{S}/\text{cm}$ | 22.2 $\mu\text{g}/\text{cm}^2$ | 222 mg/m^2 |
| 192 $\mu\text{S}/\text{cm}$ | 19.2 $\mu\text{g}/\text{cm}^2$ | 192 mg/m^2 | 223 $\mu\text{S}/\text{cm}$ | 22.3 $\mu\text{g}/\text{cm}^2$ | 223 mg/m^2 |
| 193 $\mu\text{S}/\text{cm}$ | 19.3 $\mu\text{g}/\text{cm}^2$ | 193 mg/m^2 | 224 $\mu\text{S}/\text{cm}$ | 22.4 $\mu\text{g}/\text{cm}^2$ | 224 mg/m^2 |
| 194 $\mu\text{S}/\text{cm}$ | 19.4 $\mu\text{g}/\text{cm}^2$ | 194 mg/m^2 | 225 $\mu\text{S}/\text{cm}$ | 22.5 $\mu\text{g}/\text{cm}^2$ | 225 mg/m^2 |
| 195 $\mu\text{S}/\text{cm}$ | 19.5 $\mu\text{g}/\text{cm}^2$ | 195 mg/m^2 | 226 $\mu\text{S}/\text{cm}$ | 22.6 $\mu\text{g}/\text{cm}^2$ | 226 mg/m^2 |
| 196 $\mu\text{S}/\text{cm}$ | 19.6 $\mu\text{g}/\text{cm}^2$ | 196 mg/m^2 | 227 $\mu\text{S}/\text{cm}$ | 22.7 $\mu\text{g}/\text{cm}^2$ | 227 mg/m^2 |
| 197 $\mu\text{S}/\text{cm}$ | 19.7 $\mu\text{g}/\text{cm}^2$ | 197 mg/m^2 | 228 $\mu\text{S}/\text{cm}$ | 22.8 $\mu\text{g}/\text{cm}^2$ | 228 mg/m^2 |
| 198 $\mu\text{S}/\text{cm}$ | 19.8 $\mu\text{g}/\text{cm}^2$ | 198 mg/m^2 | 229 $\mu\text{S}/\text{cm}$ | 22.9 $\mu\text{g}/\text{cm}^2$ | 229 mg/m^2 |
| 199 $\mu\text{S}/\text{cm}$ | 19.9 $\mu\text{g}/\text{cm}^2$ | 199 mg/m^2 | 230 $\mu\text{S}/\text{cm}$ | 23.0 $\mu\text{g}/\text{cm}^2$ | 230 mg/m^2 |
| 200 $\mu\text{S}/\text{cm}$ | 20.0 $\mu\text{g}/\text{cm}^2$ | 200 mg/m^2 | 231 $\mu\text{S}/\text{cm}$ | 23.1 $\mu\text{g}/\text{cm}^2$ | 231 mg/m^2 |
| 201 $\mu\text{S}/\text{cm}$ | 20.1 $\mu\text{g}/\text{cm}^2$ | 201 mg/m^2 | 232 $\mu\text{S}/\text{cm}$ | 23.2 $\mu\text{g}/\text{cm}^2$ | 232 mg/m^2 |
| 202 $\mu\text{S}/\text{cm}$ | 20.2 $\mu\text{g}/\text{cm}^2$ | 202 mg/m^2 | 233 $\mu\text{S}/\text{cm}$ | 23.3 $\mu\text{g}/\text{cm}^2$ | 233 mg/m^2 |
| 203 $\mu\text{S}/\text{cm}$ | 20.3 $\mu\text{g}/\text{cm}^2$ | 203 mg/m^2 | 234 $\mu\text{S}/\text{cm}$ | 23.4 $\mu\text{g}/\text{cm}^2$ | 234 mg/m^2 |
| 204 $\mu\text{S}/\text{cm}$ | 20.4 $\mu\text{g}/\text{cm}^2$ | 204 mg/m^2 | 235 $\mu\text{S}/\text{cm}$ | 23.5 $\mu\text{g}/\text{cm}^2$ | 235 mg/m^2 |
| 205 $\mu\text{S}/\text{cm}$ | 20.5 $\mu\text{g}/\text{cm}^2$ | 205 mg/m^2 | 236 $\mu\text{S}/\text{cm}$ | 23.6 $\mu\text{g}/\text{cm}^2$ | 236 mg/m^2 |
| 206 $\mu\text{S}/\text{cm}$ | 20.6 $\mu\text{g}/\text{cm}^2$ | 206 mg/m^2 | 237 $\mu\text{S}/\text{cm}$ | 23.7 $\mu\text{g}/\text{cm}^2$ | 237 mg/m^2 |
| 207 $\mu\text{S}/\text{cm}$ | 20.7 $\mu\text{g}/\text{cm}^2$ | 207 mg/m^2 | 238 $\mu\text{S}/\text{cm}$ | 23.8 $\mu\text{g}/\text{cm}^2$ | 238 mg/m^2 |
| 208 $\mu\text{S}/\text{cm}$ | 20.8 $\mu\text{g}/\text{cm}^2$ | 208 mg/m^2 | 239 $\mu\text{S}/\text{cm}$ | 23.9 $\mu\text{g}/\text{cm}^2$ | 239 mg/m^2 |
| 209 $\mu\text{S}/\text{cm}$ | 20.9 $\mu\text{g}/\text{cm}^2$ | 209 mg/m^2 | 240 $\mu\text{S}/\text{cm}$ | 24.0 $\mu\text{g}/\text{cm}^2$ | 240 mg/m^2 |
| 210 $\mu\text{S}/\text{cm}$ | 21.0 $\mu\text{g}/\text{cm}^2$ | 210 mg/m^2 | 241 $\mu\text{S}/\text{cm}$ | 24.1 $\mu\text{g}/\text{cm}^2$ | 241 mg/m^2 |
| 211 $\mu\text{S}/\text{cm}$ | 21.1 $\mu\text{g}/\text{cm}^2$ | 211 mg/m^2 | 242 $\mu\text{S}/\text{cm}$ | 24.2 $\mu\text{g}/\text{cm}^2$ | 242 mg/m^2 |
| 212 $\mu\text{S}/\text{cm}$ | 21.2 $\mu\text{g}/\text{cm}^2$ | 212 mg/m^2 | 243 $\mu\text{S}/\text{cm}$ | 24.3 $\mu\text{g}/\text{cm}^2$ | 243 mg/m^2 |
| 213 $\mu\text{S}/\text{cm}$ | 21.3 $\mu\text{g}/\text{cm}^2$ | 213 mg/m^2 | 244 $\mu\text{S}/\text{cm}$ | 24.4 $\mu\text{g}/\text{cm}^2$ | 244 mg/m^2 |
| 214 $\mu\text{S}/\text{cm}$ | 21.4 $\mu\text{g}/\text{cm}^2$ | 214 mg/m^2 | 245 $\mu\text{S}/\text{cm}$ | 24.5 $\mu\text{g}/\text{cm}^2$ | 245 mg/m^2 |
| 215 $\mu\text{S}/\text{cm}$ | 21.5 $\mu\text{g}/\text{cm}^2$ | 215 mg/m^2 | 246 $\mu\text{S}/\text{cm}$ | 24.6 $\mu\text{g}/\text{cm}^2$ | 246 mg/m^2 |
| 216 $\mu\text{S}/\text{cm}$ | 21.6 $\mu\text{g}/\text{cm}^2$ | 216 mg/m^2 | 247 $\mu\text{S}/\text{cm}$ | 24.7 $\mu\text{g}/\text{cm}^2$ | 247 mg/m^2 |
| 217 $\mu\text{S}/\text{cm}$ | 21.7 $\mu\text{g}/\text{cm}^2$ | 217 mg/m^2 | 248 $\mu\text{S}/\text{cm}$ | 24.8 $\mu\text{g}/\text{cm}^2$ | 248 mg/m^2 |

Ascott Analytical Equipment Limited

6-8 Gerard, Lichfield Road Industrial Estate, Tamworth, Staffordshire, B79 7UW, Great Britain

T +44 (0) 1827 318040 F +44 (0) 1827 318049 E sales@ascott-analytical.com W www.ascottshop.com