

# pH Meter

Ascott Ref No ACC11



(Picture may differ from actual item supplied)

**Ascott Analytical Equipment Limited,**  
6 - 8 Gerard, Lichfield Road Industrial Estate,  
Tamworth, Staffs. B79 7UW Great Britain  
Telephone: +44 (0)1827 318040 Fax: +44 (0)1827 318049  
E-Mail: [info@ascott-analytical.com](mailto:info@ascott-analytical.com) Website: [www.ascott-analytical.com](http://www.ascott-analytical.com)

## Description

Hand held, battery powered digital pH meter, for the measurement of pH of the salt fog fallout over the range 0-14 with a resolution of 0.01 pH. The meter is provided with manual temperature compensation, slope and calibration controls. Supplied complete with buffer solutions for calibration.

## Installation

Not applicable.

## Operation

See the manufacturer's manual.

## Maintenance

See the manufacturer's manual.

<i>Issue</i>	<i>Date First Issued</i>	<i>Description Of Amendments</i>
A	December 2010	Original Issue
B	March 2015	Layout Updated