

pH Meter



Ascott Ref No ACC11

(Picture may differ from actual item supplied)

Ascott Analytical Equipment Limited,
6 - 8 Gerard, Lichfield Road Industrial Estate,
Tamworth, Staffs. B79 7UW Great Britain
Telephone: +44 (0)1827 318040 Fax: +44 (0)1827 318049
E-Mail: info@ascott-analytical.com Website: www.ascott-analytical.com

Description

Hand held, battery powered digital pH meter, for the measurement of pH of the salt fog fallout over the range 0-14 with a resolution of 0.01 pH. The meter is provided with manual temperature compensation, slope and calibration controls. Supplied complete with buffer solutions for calibration.

Installation

Not applicable.

Operation

See the manufacturer's manual.

Maintenance

See the manufacturer's manual.

<i>Issue</i>	<i>Date First Issued</i>	<i>Description Of Amendments</i>
A	December 2010	Original Issue
B	March 2015	Layout Updated
C	October 2025	Updated image